

Service Information Letter 01_2010

1 FW update for the DA-100 NT

1.1 Initial position

MBV AG is continuously developing their instruments to provide service for our customers and to improve the quality of our products.

1.2 MBV AG action

MBV AG has released a new firmware version of the DA-100NT which doesn't have any effect to the measurement results of the instrument.

1.3 FW history

V.1.7 (04.09.2007)
First released version.

V.1.8 (10.12.2007)
A new shut off routine which prevents discharging of the battery.

V.1.9 (14.01.2008)
*Factory settings: "Temperature and pressure sensor are programeble."
Only compatible with PC-SW ≥ 1.4 .*

1.4 Notes

The DA-100 NT doesn't has to be calibrated after a firmware update.

1.5 Who will proceed the update?

The Firmware update has to be done by MBV AG, Switzerland.

Stäfa, 04.02.2010



MBV AG
Hans Zingre
CEO



MBV AG
Cedomir Krndija
Production Manager / QM