

Service Information Letter 02_2010

1 Automatical calibration of the MAS-100CG Ex compressed gas air sampler

1.1 Initial position

MBV AG is continuously developing their instruments to provide service for our customers and to improve the quality of our products.

The calibration of the MAS-100CG Ex unit is done by comparing its flow measurement results with a calibrated 'Radial Blade Turbine Gas Meter' TRZ/50/60 of GWF in Switzerland (www.gwf.ch).

The newest MBV AG innovation is the automatical calibration of the MAS-100 CG Ex units with the major advantage that the results will be reached automatically and the calibration equipment is based on the newest technologies.

1.2 MBV AG action

For this reason the firmware of the instruments will be updated to the latest revision 3.79 free of charge.

1.3 Notes

The functionality of the instrument and all relevant parameters are not affected with this firmware update.

Stäfa, 10.03.2010



MBV AG
Hans Zingre
CEO



MBV AG
Cedomir Krndija
Production Manager / QM