Ref. Certif. No.

CH1-00014

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Microbial Air Sampler
Name and address of the applicant	MBV AG, Industriestrasse 9, 8712 Stäfa, SWITZERLAND
Name and address of the manufacturer	MBV AG, Industriestrasse 9, 8712 Stäfa, SWITZERLAND
Name and address of the factory	Femron AG, Stadtweg 24, 8245 Feuerthalen, SWITZERLAND
Note: *When more than one factory, please report on page 2*	☐ Additional Information on page 2
Ratings and principal characteristics	Input: 5 V ⎓ 3.0 A; 20 V ⎓ 1.5 A; Battery: Li-on 7.38 V ⎓ 9.6 Ah; 70.85 W
Trademark / Brand (if any)	mBV
Customer's Testing Facility (CTF) Stage used	./.
Model / Type Ref.	MAS-100 Sirius® and MAS-100 Sirius® Ex
Additional information (if necessary may also be reported on page 2)	National Differences: EU Group Differences, CA, JP, US ☐ Additional Information on page 2
A sample of the product was tested and found to be in conformity with	IEC 61010-1:2010 IEC 61010-1:2010/AMD1:2016
As shown in the Test Report Ref. No. which forms part of this Certificate	EMCKP5298.1A

This CB Test Certificate is issued by the National Certification Body

emc testcenter.zürich.ag

EMC-Testcenter AG
Moosäckerstrasse 77
8105 Regensdorf
SWITZERLAND

EMC-Testcenter AG Certification Body

Date: 2025-03-05

Signature: Ulrike HIEGEMANN