

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Settle plate changer

Name and address of the applicant

MBV AG, Industriestrasse 9, 8712 Stäfa,
SWITZERLAND

Name and address of the manufacturer

MBV AG, Industriestrasse 9, 8712 Stäfa,
SWITZERLAND

Name and address of the factory

Femron AG, Stadtweg 24, 8245 Feuerthalen,
SWITZERLAND*Note: When more than one factory, please report on page 2*☐ Additional Information on page 2

Ratings and principal characteristics

24 V DC, ± 2.4 V, 2 A max.

Trademark / Brand (if any)



Customer's Testing Facility (CTF) Stage used

./.

Model / Type Ref.

MAS-100 Libra®

Additional information (if necessary may also be reported on page 2)

National Differences: EU Group Differences, CA, US, KR

☒ Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 61326-1:2020


As shown in the Test Report Ref. No. which forms part of this Certificate

EMCKP6085.1C, EMCKP6085.1D


This CB Test Certificate is issued by the National Certification Body



emc testcenter.zürich.ag

EMC-Testcenter AG
Moosackerstrasse 77
8105 Regensdorf
SWITZERLAND

Signature: Klaus SCHMIDT



Date: 2025-02-24

Additional Standards:

IEC 61326-1:2012 (ed.2)
CISPR 11:2015, Class A
CISPR 11:2015/AMD1:2016, Class A
CISPR 11:2015/AMD2:2019, Class A
CISPR 32:2015, Class A
CISPR 32:2015/AMD1:2019, Class A
IEC 61000-3-2:2014 (ed.4)
IEC 61000-3-2:2018 (ed.5)
IEC 61000-3-2:2018/AMD1:2020
IEC 61000-3-3:2013 (ed.3)
IEC 61000-3-3:2013/AMD1:2017
IEC 61000-3-3:2013/AMD2:2021

Additional information (if necessary)