



GlobTek[®], Inc.

医疗、军工、资讯、车载、特种电源

ISO 9001:2008
REGISTERED

PSE Products Inspection Record

PART#: TR9K12700L6PNR6B3122 L/R#: 8003484

Name of Electrical Appliances: DC Power Supply

Model Name: GT-46600-6524-72

Classification: Refer to Appendix A

Construction, Material and Performance: Refer to Appendix A

Date/Month/Year of the Test:

13-JULY-2020

Date code: 2720

Place of the Test:

Test conducted by: 張華花 2020.7.13

Number of Electrical Appliances tested: 540 pcs

Serial Number: N/A

Test Method: Chroma Automatic Test Equipment, Hi-Pot Tester

Test Result:

Input Current < 1.5A @ 100VAC at full load

Visible Inspection

Dielectric Strength

Functioning Test

I/P to O/P
AC: 3000V/10mA/60S

DC: 500V/100mA/2S

Pass except below
mentioned

Pass except below
mentioned

Pass except below mentioned

540 pcs - PASS

540 pcs - PASS

540 pcs - PASS

Serial No. of Failed products

N/A

Nil

Nil

Nil

Nil

Nil

Nil

Nil

Nil

Nil

This record shall be kept for three years from the day test conducted.

Approved: _____

